

# Transparent Tip Protectors

<b>Product Code Number(s)</b>	10001, 10002, 10003, 10004, 10005, 10006, 10007, 10008, 10009, 10010
<b>Intended Use</b>	Prevents moisture from being trapped between the protector and the instrument. Tip Protectors are also vented – insuring that moisture quickly evaporates during the sterilization cycle. The see through design means tips are easily visible.
<b>Range of Applications for Product</b>	Small to large tipped instruments.
<b>Key specifications of product</b>	<ul style="list-style-type: none"> <li>• Compatible with steam and EO sterilization.</li> <li>• Ribs which promote sterilant circulation, and reduce chance for sticking to instruments.</li> <li>• Vent holes to promote sterilant circulation.</li> <li>• Translucent material allows for view of instrument tip.</li> <li>• Available in 10 sizes.</li> <li>• Manufactured of heat stable TPE.</li> <li>• Single-use</li> </ul>

## Shipping & Storage

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	100 per box
<b>Shelf Life</b>	

## Instructions for Using Product

<b>Description of Use (s)</b>	To be placed on surgical instruments such as: tweezers, iris scissors, mosquito clamp, osteotome, kochers, male scissors and other surgical instruments during sterilization to protect the delicate and sharp tips.
<b>Preparation for Use</b>	Fit each instrument with the appropriate size tip.
<b>Diagrams (drawings, pictures):</b>	
<b>Steps for Use of Product</b>	<ul style="list-style-type: none"> <li>• Place protector tip on appropriate sized instrument prior to sterilization.</li> <li>• Package instruments in sterile packaging device (e.g., pouch, wrapped tray, closed container).</li> <li>• Place package in sterilizer (steam or EO).</li> <li>• Dispose of after use in compliance with facility guidelines.</li> </ul>
<b>Interpretation of Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	
<b>Disposal</b>	Dispose of in compliance with facility guidelines.

## Reprocessing Instructions

<b>Point of use:</b>	
<b>Preparation for decontamination:</b>	
<b>Disassembly Instructions:</b>	
<b>Cleaning – Manual:</b>	
<b>Cleaning – Automated:</b>	
<b>Disinfection:</b>	
<b>Drying:</b>	
<b>Maintenance, inspection, and testing:</b>	
<b>Reassembly Instructions:</b>	
<b>Packaging:</b>	
<b>Sterilization:</b>	<ul style="list-style-type: none"> <li>• Compatible with steam sterilization. Recommended cycle is 4 minutes at 270°F or 3 minutes at 275°F in a pre-vac sterilizer.</li> <li>• Material compatibility with EO sterilization.</li> </ul>